

Limited Liability Company
“Center for quality control BIOLIFE”

115478, Moscow, Kashirskoe Hwy, 24, bldg. 2. Tel: 8 (495)781-20-42, 8 (495) 781-20-43
Accreditation certificate № RA.RU.21.ЦК01 dated 22.07.2015

APPROVED
The Head of Testing Laboratory
Of CQC Biolife, LLC
Razlivayeva O.V.

<signature>
“6” June 2019

**Round Seal: /limited liability company “Center for quality control BIOLIFE”,
PSRN 1147746879280, MOSCOW/**

PROTOCOL
TOXICOLOGY TESTING OF MEDICAL DEVICES (MATERIALS)
FOR THE PURPOSE TO ESTABLISH THEIR SAFETY

№ 147-04П DATED 6TH OF JUNE, 2019

Total pages: 4

| | |
|---|---|
| Name of device (material) | HBIS Endoprosthesis of synovial fluid NOLTREX, Technical Specification file (TU) 9398-001-52820385-2015 (see Annex on 1 page). |
| Manufacturer | RESEARCH CENTER BIOFORM, LLC located 142784, Moscow, s.Moskovskiy, Kievskoe Hwy, 22 nd km, house 4, bldg. 2, floor 5, block G Manufacturing address: 142784, Moscow, s.Moskovskiy, Kievskoe Hwy, 22 nd km, house 4, bldg. 2, floor 5, block G |
| Applicant | RESEARCH CENTER BIOFORM, LLC 142784, Moscow, s.Moskovskiy, Kievskoe Hwy, 22 nd km, house 4, bldg. 2, floor 5, block G |
| Date of sample presented for testing | 23.05.2019 |
| Risk class | 3 |
| No of application for testing | №1 |

The testing is performed in accordance with (list of standards and other normative documents):

GOST R ISO 10993-2-2009

GOST ISO 10993-1-2011, 10993-5-2011, 10993-6-2011, 10993-7-2011, 10993-9-2015, 10993-10-2011, 10993-11-2011, 10993-12-2015 “Biological evaluation of medical devices”

GOST R ISO 52770-2016 “Medical devices. Safety requirements. Methods of sanitary-chemical and toxicological tests”

GOST 31214-2016 “Medical devices. Requirements for samples and documentation presented for toxicological tests, sanitary and chemical analyses, tests for sterility and pyrogenicity”

Page 1 of 4

Protocol №147-04П of 17.06.2019

GOST 31209-2003 Containers for blood and its components. Requirements for chemical and biological safety and methods of testing., article 5.3

GOST 31576-2012 Evaluation of biological hazard of medical dental materials and articles. Classification and sampling.

TESTING RESULTS

| № | Parameter | Permissible values | Results | Conclusion |
|------|--|---|---|------------|
| 1.1 | Reducing impurities | less than 1,0ml. (0,02H sodium thiosulfate solution) | 0.22 – 0.78 | complies |
| 1.2 | Change in pH | less than 1,0 | 0.36-0.87 | complies |
| 1.3 | Ultraviolet absorption | less than 0,3 (range 230-260nm.) | 0.0411-0.1049 | complies |
| 1.4 | Concentration of hazard impurities and heavy metals in aqueous extracts | | | |
| | Zinc | 1.000 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Iron | 0.030 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Chromium | 0.100 mg/l | Below determination limit of 0.0025 mg/l | complies |
| | Copper | 1.000 mg/l | Below determination limit of 0.0005 mg/l | complies |
| | Aluminum | 0.500 mg/l | Below determination limit of 0.002 mg/l | complies |
| | Silver | 0.050 mg/l | Below determination limit of 0.00005 mg/l | complies |
| | Formaldehyde | 0.100 mg/l | Below determination limit of 0.01 mg/l | complies |
| | Acetaldehyde | 0.200 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Acetone | 0.100 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Methanol | 0.200 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Propyl alcohol | 0.100 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Isopropyl alcohol | 0.100 mg/l | Below determination limit of 0.005 mg/l | complies |
| | Acrylonitrile | 0.020 mg/l | Below determination limit of 0.002 mg/l | complies |
| 2 | Toxicology test results | | | |
| 2.1 | <i>Irritating effect on the skin and mucous membrane of animals in points</i> | | | |
| | Skin | 0 | 0 | complies |
| | Conjunctiva | 0 | 0 | complies |
| 2.2. | Cytotoxicity | | | |
| | Cytotoxicity with culture of fibroblasts of mice NIH/3T3 | Permissible value \leq 30% | 0-5% | complies |
| 2.3. | <i>Acute toxicity in white mice at intra-abdominal introduction</i> | | | |
| | Deaths | none | none | complies |
| | Clinical symptoms of intoxication | none | none | complies |
| | Macroscopic changes of organs | none | none | complies |
| | Weight coefficient of visceras (observed changes) | none | none | complies |

**Conclusions as per results of testing:
Tested samples comply with the requirements for medical devices, do not cause negative effects on biological objects.**

CONCLUSION

HBIS Endoprosthesis of synovial fluid NOLTREX, technical file specification TU 9398-001-52820385-2015 (see Annex 1 on 1 page), manufactured by RC BIOFORM, LLC complies with normative documentation as per toxicity parameters for medical devices.

GOST R ISO 10993-2-2009

GOST ISO 10993-1-2011, 10993-5-2011, 10993-6-2011, 10993-7-2011, 10993-9-2015, 10993-10-2011, 10993-11-2011, 10993-12-2015 “Biological evaluation of medical devices”

GOST R ISO 52770-2016 “Medical devices. Safety requirements. Methods of sanitary-chemical and toxicological tests”

GOST 31214-2016 “Medical devices. Requirements for samples and documentation presented for toxicological tests, sanitary and chemical analyses, tests for sterility and pyrogenicity”

GOST 31209-2003 Containers for blood and its components. Requirements for chemical and biological safety and methods of testing., article 5.3

GOST 31576-2012 Evaluation of biological hazard of medical dental materials and articles. Classification and sampling.

The testing was done in accordance with normative documentation for such type of tests and using verified laboratory equipment;

Spectrophotometer LEKI SS2109UV; atomic absorption spectrometer ICE 3500; gas chromatograph 6000 Vega Series 2GC; gas chromatograph Crystal 2000M; gas chromatograph Chromatech-Crystall 5000; pH-meter pH-150MM; microscope Standard 25

RECOMMENDED for use as per toxicity parameter.

Note: the present protocol is applicable to the tested samples only, that were presented by the applicant. The partial copy or full reprint of this protocol is prohibited.

Responsible for testing:

A.A. Zabegaeva

Head of medical device testing department of CQC Biolife, LLC

K.V. Taraban

Specialist of medical device testing department of CQC Biolife, LLC

Round Seal: /limited liability company “Center for quality control BIOLIFE”,
PSRN 1147746879280, MOSCOW/

Annex 1 to the toxicology testing protocol №147-04П

HBIS Endoprosthesis of synovial fluid NOLTREX as per technical specification file TU 9398-001-52820385-2015, device variations:

1. – NOLTREX 2.5
2. – NOLTREX 3.0
3. – NOLTREX 5.0

Responsible for testing:

A.A. Zabegaeva

Head of medical device testing department of CQC Biolife, LLC

K.V. Taraban

Specialist of medical device testing department of CQC Biolife, LLC

**Round Seal: /limited liability company “Center for quality control BIOLIFE”,
PSRN 1147746879280, MOSCOW/**